

**PROCUREMENT OF GOODS
UNDER
SHOPPING PROCEDURES**

(For Contracts valued less than US \$ 30,000 each)

Dt: 21-02-2020

REF: NCSCM/PROC/Autoclave/2020/G-82

**INVITATION FOR QUOTATIONS FOR SUPPLY OF
GOODS UNDER SHOPPING PROCEDURES**

To

SUB: INVITATION FOR QUOTATIONS FOR SUPPLY & INSTALLATION OF
AUTOCLAVE

Dear Sir,

1. You are invited to submit your most competitive quotation for the supply of following goods:

Brief Description of the Goods	Specifications	Quantity	Delivery Period	Place of Delivery	Installation Requirement if any
SUPPLY & INSTALLATION OF AUTOCLAVE	ANNEX – I	2 Nos	15 Days from the Date of Purchase Order	NCSCM, Chennai	YES

2. Government of India has received a credit from the International Development Association (IDA) towards the cost of the Project and intends to apply part of the proceeds of this credit to eligible payments under the contract for which this invitation for quotations is issued.

3. Documents Establishing Bidder's Eligibility and Qualifications:

3.1 The Bidder shall furnish, as part of its quotation, following documents to establish the Bidder's eligibility to quote and its qualifications to perform the Contract if its quotation is accepted -

- (a) The legal status, place of registration, place of business of the company, or firm or partnership;
- (b) Copy of GST Registration Certificate and TAN;
- (c) PAN with self-attested copy of PAN Card;
- (d) Must have supplied, installed and commissioned satisfactorily at least 100% of the quantity similar to the type specified in the "Schedule of Requirements" in any one of the past three financial years i.e. 2016-17 to 2018-19.
- (e) Must have a minimum average annual turnover of Rs.4 lakh during the last three financial years i.e., 2016-17 to 2018-19.

The bidders should furnish the above information along with prescribed format attached. (Attachment 1)

4. **Bid Price**

- a) The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialing, dating and re-writing.
- b) All duties, taxes and other levies payable on the raw materials and components shall be included in the total price.
- c) GST shall be shown separately.
- d) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- e) The Prices shall be quoted in Indian Rupees only.

5. Each bidder shall submit only one quotation. Bidder shall not contact other Bidders in matters relating to this Quotation.

6. **Validity of Quotation**

Quotation shall remain valid for a period not less than 90 days after the deadline date specified for submission.

7. **Evaluation of Quotations**

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

- (a) Are properly signed;
- (b) Conform to the terms and conditions, and specifications and
- (c) Has submitted all the documents asked for in 3 above

The Quotations would be evaluated for all the items together.

GST shall not be taken into account in evaluation.

8. **Documents Establishing Goods' Conformity to Bidding Documents**

The documentary evidence of conformity of the goods and services to the bidding documents may be in the form of literature or drawings or any other documentary proof.

9. **Award of contract**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

9.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.

- 9.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the supply order.
10. Payment shall be made immediately after delivery of the goods.
11. Normal commercial warranty/ guarantee shall be applicable to the supplied goods.
12. You are requested to provide your offer latest by 03-03-2020 hours on 11.00 AM.
13. We look forward to receiving your quotations and thank you for your interest in this project.

**Sd/-
Director
National Centre for Sustainable Coastal Management
Anna University Campus
Chennai – 600 025**

FORMAT OF QUOTATION *

Item No.	Description of the goods	Quantity	Unit	Quoted unit price for delivery up to destination	Unit Cost of Incidental Services (Rs)	Total Unit cost (Rs.) [Col(5+6)]		Amount (Rs.) [Col(3x7)]	GST (%)
						In Figures	In words		
1	2	3	4	5	6	7	8	9	10

Gross Total Cost: Rs.

We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs.(amount in figures) (Rs. amount in words) within the period specified in the Invitation for Quotations.

We also confirm that the normal commercial warranty/guarantee of months shall apply to the offered goods.

We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in bribery or collusive agreements with competitors.

Signature of Supplier

TECHNICAL SPECIFICATION

Technical Specification for the Autoclave

Quantity Required: 2 Numbers

1	Technical specification for Autoclave (15L): 2 Nos.
	<p>Chamber Volume (Litre) ~15 Litre Power Consumption (KW): 1 KW</p> <ul style="list-style-type: none"> ○ Complete body of stainless steel of grade SS 304. ○ Capable of sterilization at 121°C and 134°C with pre-selectable sterilization time. ○ Dry cycle with selectable time period with alarm to indicate end of process. ○ Compatible sterilization tray should be given with the autoclave. ○ Visual indicator provided by Red LED to indicate process status. Safety valve to release excess pressure to prevent accident. ○ Alphanumeric display
2	Technical specification for Autoclave (30L): 2 Nos.
	<p>Chamber Volume (Litre) ~30 Litre Power Consumption (KW): 1 KW</p> <ul style="list-style-type: none"> ○ Complete body of stainless steel of grade SS 304. ○ Capable of sterilization at 121°C and 134°C with pre-selectable sterilization time. ○ Pre-vacuum cycle to remove air from the chamber. Water tank inbuilt with automatic water level sensors. ○ Automatic pressure door locking system. Automatic process termination in case, the door is opened accidentally. ○ Dry cycle with selectable time period with Audio-visual alarm to indicate end of process. ○ Compatible sterilization tray should be given with the autoclave. ○ Visual indicator provided by Red LED to indicate process status. Safety valve to release excess pressure to prevent accident. ○ Alphanumeric display to set the protocol with soft keypad.